

List of CAMAC boards used @ WHT, detailed status of the spares available.

Last update of this list: 20/10/2005

Camac board type	Use at WHT	S/N	Latest test	Status if non OK
Hytec 1386 Camac dataway controller	Branch 0 Crate 0, x1	N 117 "Pass-thru mode"	06/11/2003	
		N 118 "Pass-thru mode"	06/11/2003	
		N 101 "JKT unit, non pass-thru"	06/11/2003	
		N 102 "JKT unit, non pass-thru"	06/11/2003	
PR2402 // input register (read UTC from time service)	B0 C0 x1, B6 C3 x2	RGO 6724	18/09/2003	
		RGO 6847	18/09/2003	
		RGO 6844	18/09/2003	
		RGO 6843	18/09/2003	
		RGO 6842	18/09/2003	
LP34 millisecond generator LP37	B0 C0 x1	RGO 6849	17/09/2003	pb O/P B
	B6 C1 x1	94/02	11/12/2003	
RGO 32bit counters (test is to read one HA Ferranti encoder, compare with other 2)	B6 C1 x4, B6 C2 x2	Issue 3 S/N 01	22/09/2003	pb of synchro?
		Issue 2 S/N 93/004	22/09/2003	
		Issue 2 S/N 93/003	22/09/2003	
there are 6 non used counters at WHT, they were intended for the tape encoders.				
ED12 Pulse Generator (generates LAMs for the TCS)	B0 C0, x1	no S/N	06/11/2003	
		no S/N	06/11/2003	
ED13	B6 C1 x1		no spare	
ED11	B6 C2 x1		no spare	
BR CPR 4-1 Branch coupler	B0 C0	S/N 027	New; replaced 06/2005 by Hytec for unit S/N 141 used in the CAMAC test station	
		S/N 1622		
BR CPR 3 Branch coupler	old version of CPR 4. Use as spare?	S/N 57	07/05/1993	
BT 6502 Branch terminator	B6 C3 x1	S/N 98 @ int	29/09/1993	
		RGO 2784 @wht	no date	
CC 2089 Crate controller	B6 C1 x1, B6 C2 x1, B6 C3 x1	RGO 1947	11/02/1993	
ADC 1232 range selection depends on the generation of board: 1232, -1, -2	B6 C3 x2. Need to set range 0-1V.	no S/N	Hytec 1232-2	14/09/1994
		S/N 8	Hytec 1232	04/08/1994
DTM4	Only used for visual tests	S/N 553	28/12/1995	

	plenty of spares, no need to test	S/N 885 S/N 1859	29/09/1993 no date		
OD48	B6 C3, x2	S/N 10500B	14/03/2001		
OR48	B6 C1, x3	S/N 10510	19/04/2005	Bad altitude (Reg A) rate@ wht	
OD2411	cannot be tested @ INT	B6 C1 x1	S/N 016	no date	
450-4	only used at wht	B6 C2 x7 B6 C3 x1	1502 "links for WHT"	27/03/2000	"not to be used"
see http://www.ing.iac.es/~eng/electronics/legacy/steve_magee/camac/index.htm	for description of links settings	1537	09/02/2000		
		1535	09/02/2000		
		1903	09/02/2000		
		1516	16/06/2005	repaired by Hytec	
		1532	16/06/2005	repaired by Hytec	
		1534	16/06/2005	repaired by Hytec	
		no S/N, id.	22/09/2001	at INT elec. Lab	

Other types of boards only used at INT

	used in:			
OD 2407 output driver	B4 C3 x3			
PR 2403 parallel data input register	B4 C3 x1	S/N 0101	16/06/2005	repaired by Hytec
Multidac 9085 or Hytec 640VSL	B4 C3 x1			
3340 RS232 interface	B4 C3 x3			

Other types of boards only used at JKT

DBE 6501 differential branch extender	B4 C1 x1, B4 C2 x1
--	--------------------